

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00031	Cert. No.: 00001
Sequential File No.:	00031	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 12/15/2005
		Soln. No.: 00004
		File Date: 12/15/2005
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0631

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:51S		
Control 1 EC	0.099%	09:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		
Control 2 EC	0.098%	09:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54S		
Control 3 EC	0.098%	09:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: _____



Badge No.: 513

Date: 12/15/2005

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	Calib. No.:	00003
Calibration File No.:	00008	Cert. No.:	00001
Certification File No.:	00009	Lin. No.:	00001
Linearity File No.:	00010	Soln. No.:	00003
Solution File No.:	00017	Calib. Date:	10/18/2005
Sequential File No.:	00017	Cert. Date:	10/18/2005
		Lin. Date:	10/18/2005
		Soln. Date:	11/17/2005
		File Date:	11/17/2005
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0329
Solution Control Lot:	05F019	Expires:	06/03/2007
		Bottle No.:	0626

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:23S		
Control 1 EC	0.099%	15:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		
Control 2 EC	0.098%	15:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26S		
Control 3 EC	0.098%	15:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: 

Badge No.: 513

Date: 11/17/2005

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00011	Cert. No.: 00001
Sequential File No.:	00011	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 10/18/2005
		Soln. No.: 00002
		File Date: 10/18/2005
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0642

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:14D		
Control 1 EC	0.100%	11:15D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:15D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 2 EC	0.099%	11:16D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:16D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.099%	11:17D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:17D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR. I WILL M Cross #5342

Badge No.: 5342

Date: 10/18/2005