

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: WASHINGTON TOWNSHIP PD
 Calibration File No.: 00147
 Certification File No.: 00148
 Linearity File No.: 00149
 Solution File No.: 00183
 Sequential File No.: 00183

Serial No.: ARWF-0369
 Calib. Date: 10/13/2006
 Cert. Date: 10/13/2006
 Lin. Date: 10/13/2006
 Soln. Date: 12/30/2006
 File Date: 12/30/2006
 Calib. No.: 00004
 Cert. No.: 00002
 Lin. No.: 00002
 Soln. No.: 00027

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 05L022

Model No.: CU-34
 Serial No.: DDWJ S3-0323
 Expires: 12/15/2007
 Bottle No.: 0497


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:44S		
Control 1 EC	0.100%	16:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 2 EC	0.099%	16:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46S		
Control 3 EC	0.098%	16:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NITTI First Name: CHRISTOPHER MI: A

Signature:  Badge No.: 517
 Date: 12/30/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00147	Calib. Date: 10/13/2006
Certification File No.:	00148	Calib. No.: 00004
Linearity File No.:	00149	Cert. Date: 10/13/2006
Solution File No.:	00171	Cert. No.: 00002
Sequential File No.:	00171	Lin. Date: 10/13/2006
		Lin. No.: 00002
		Soln. Date: 12/15/2006
		Soln. No.: 00026
		File Date: 12/15/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05L022	Expires: 12/15/2007
		Bottle No.: 0490

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:28S		
Control 1 EC	0.100%	14:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:29S		
Control 2 EC	0.098%	14:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		
Control 3 EC	0.097%	14:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: _____



Badge No.: 204

Date: 12/15/2006

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00147	Calib. Date: 10/13/2006
Certification File No.:	00148	Calib. No.: 00004
Linearity File No.:	00149	Cert. Date: 10/13/2006
Solution File No.:	00163	Cert. No.: 00002
Sequential File No.:	00163	Lin. Date: 10/13/2006
		Lin. No.: 00002
		Soln. Date: 11/25/2006
		Soln. No.: 00025
		File Date: 11/25/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05L022	Expires: 12/15/2007
		Bottle No.: 0485

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:08S		
Control 1 EC	0.099%	06:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 2 EC	0.097%	06:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		
Control 3 EC	0.097%	06:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NITTI	First Name: CHRISTOPHER	MI: A
Signature: 	Badge No.: 517	Date: 11/25/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00147	Calib. Date: 10/13/2006
Certification File No.:	00148	Calib. No.: 00004
Linearity File No.:	00149	Cert. Date: 10/13/2006
Solution File No.:	00158	Cert. No.: 00002
Sequential File No.:	00158	Lin. Date: 10/13/2006
		Lin. No.: 00002
		Soln. Date: 11/08/2006
		Soln. No.: 00024
		File Date: 11/08/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05L022	Expires: 12/15/2007
		Bottle No.: 0493

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:39S		
Control 1 EC	0.099%	15:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41S		
Control 2 EC	0.097%	15:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:42S		
Control 3 EC	0.097%	15:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: _____



Badge No.: 204

Date: 11/08/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0369
Location: WASHINGTON TOWNSHIP PD
Calibration File No.: 00147 Calib. Date: 10/13/2006 Calib. No.: 00004
Certification File No.: 00148 Cert. Date: 10/13/2006 Cert. No.: 00002
Linearity File No.: 00149 Lin. Date: 10/13/2006 Lin. No.: 00002
Solution File No.: 00153 Soln. Date: 10/29/2006 Soln. No.: 00022
Sequential File No.: 00153 File Date: 10/29/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0323
Control Solution %: 0.100% Expires: 12/15/2007
Solution Control Lot: 05L022 Bottle No.: 0222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:33S		
Control 1 EC	0.100%	18:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:34S		
Control 2 EC	0.098%	18:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:36S		
Control 3 EC	0.097%	18:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:37S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NITTI

First Name: CHRISTOPHER

MI: A

Signature: 

Badge No.: 517

Date: 10/29/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00144	Cert. No.: 00001
Sequential File No.:	00144	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 10/03/2006
		Soln. No.: 00019
		File Date: 10/03/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0647

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:23D		
Control 1 EC	0.099%	23:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:24D		
Control 2 EC	0.098%	23:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:25D		
Control 3 EC	0.098%	23:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: _____



Badge No.: 204

Date: 10/03/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00147	Calib. Date: 10/13/2006
Certification File No.:	00148	Calib. No.: 00004
Linearity File No.:	00149	Cert. Date: 10/13/2006
Solution File No.:	00151	Cert. No.: 00002
Sequential File No.:	00151	Lin. Date: 10/13/2006
		Lin. No.: 00002
		Soln. Date: 10/13/2006
		Soln. No.: 00021
		File Date: 10/13/2006

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0323
Control Solution %:	0.100%			Expires:	12/15/2007
Solution Control Lot:	05L022			Bottle No.:	0494

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:58D		
Control 1 EC	0.100%	15:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59D		
Control 2 EC	0.098%	16:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00D		
Control 3 EC	0.098%	16:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWAP2-134

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: *TH. II Thomas J. Snyder #5792*

Badge No.: 5792

Date: 10/13/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00147	Calib. Date: 10/13/2006
Certification File No.:	00148	Cert. No.: 00004
Linearity File No.:	00149	Cert. Date: 10/13/2006
Solution File No.:	00150	Lin. Date: 10/13/2006
Sequential File No.:	00150	Soln. Date: 10/13/2006
		Lin. No.: 00002
		Soln. No.: 00020
		File Date: 10/13/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0648

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:13D		
Control 1 EC	0.098%	14:13D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	14:13D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:14D		
Control 2 EC	0.000%	14:15D	33.8°C	CONTROL TEST FAILED
Control 2 IR	0.000%	14:15D	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:16D		
Control 3 EC	0.100%	14:17D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:17D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: *THOMAS J. SNYDER #5792*

Badge No.: 5792

Date: 10/13/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00138	Cert. No.: 00001
Sequential File No.:	00138	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 09/09/2006
		Soln. No.: 00018
		File Date: 09/09/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0323
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0643

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:06D		
Control 1 EC	0.099%	15:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 2 EC	0.097%	15:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		
Control 3 EC	0.098%	15:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: _____



Badge No.: 513

Date: 09/09/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0369
Location:	WASHINGTON TOWNSHIP PD		
Calibration File No.:	00008	Calib. Date:	10/18/2005
Certification File No.:	00009	Cert. Date:	10/18/2005
Linearity File No.:	00010	Lin. Date:	10/18/2005
Solution File No.:	00132	Soln. Date:	08/25/2006
Sequential File No.:	00132	File Date:	08/25/2006
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0323
Solution Control Lot:	05F019	Expires:	06/03/2007
		Bottle No.:	0641

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:51D		
Control 1 EC	0.099%	10:52D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:52D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 2 EC	0.097%	10:53D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:53D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 3 EC	0.097%	10:54D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:54D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWRP2-134 WR

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: *TPR. I. WILLIAM M. CROSS #5342*

Badge No.: 5342

Date: 08/25/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Cert. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00131	Lin. Date: 10/18/2005
Sequential File No.:	00131	Soln. Date: 08/12/2006
		File Date: 08/12/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0646

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:08D		
Control 1 EC	0.100%	19:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10D		
Control 2 EC	0.097%	19:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:11D		
Control 3 EC	0.097%	19:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: _____



Badge No.: 204

Date: 08/12/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Cert. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00127	Lin. Date: 10/18/2005
Sequential File No.:	00127	Soln. Date: 07/12/2006
		Lin. No.: 00001
		Soln. No.: 00015
		File Date: 07/12/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0640

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:48D		
Control 1 EC	0.100%	16:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49D		
Control 2 EC	0.098%	16:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:51D		
Control 3 EC	0.098%	16:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: _____



Badge No.: 513

Date: 07/12/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Cert. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00119	Lin. Date: 10/18/2005
Sequential File No.:	00119	Soln. Date: 06/12/2006
		Lin. No.: 00001
		Soln. No.: 00013
		File Date: 06/12/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0639

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:00D		
Control 1 EC	0.099%	13:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01D		
Control 2 EC	0.097%	13:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 3 EC	0.097%	13:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: 

Badge No.: 513

Date: 06/12/2006

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00108	Cert. No.: 00001
Sequential File No.:	00108	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 05/26/2006
		Soln. No.: 00012
		File Date: 05/26/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0634

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:01D		
Control 1 EC	0.099%	13:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		
Control 2 EC	0.097%	13:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:03D		
Control 3 EC	0.097%	13:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: _____



Badge No.: 513

Date: 05/26/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00099	Cert. No.: 00001
Sequential File No.:	00099	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 04/25/2006
		Soln. No.: 00011
		File Date: 04/25/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0638

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:08D		
Control 1 EC	0.100%	18:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:09D		
Control 2 EC	0.098%	18:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:10D		
Control 3 EC	0.097%	18:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: 

Badge No.: 513

Date: 04/25/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00095	Cert. No.: 00001
Sequential File No.:	00095	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 04/21/2006
		Soln. No.: 00010
		File Date: 04/21/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0636

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:09D		
Control 1 EC	0.099%	18:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:11D		
Control 2 EC	0.097%	18:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12D		
Control 3 EC	0.097%	18:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:13D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: 

Badge No.: 513

Date: 04/21/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00086	Cert. No.: 00001
Sequential File No.:	00086	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 03/24/2006
		Soln. No.: 00009
		File Date: 03/24/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 6280

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:32S		
Control 1 EC	0.100%	15:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33S		
Control 2 EC	0.099%	15:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34S		
Control 3 EC	0.098%	15:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature:



Badge No.: 204

Date: 03/24/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00079	Cert. No.: 00001
Sequential File No.:	00079	Lin. Date: 10/18/2005
		Lin. No.: 00001
		Soln. Date: 03/11/2006
		Soln. No.: 00008
		File Date: 03/11/2006
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0637

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:26S		
Control 1 EC	0.099%	11:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 2 EC	0.099%	11:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		
Control 3 EC	0.097%	11:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: 

Badge No.: 204

Date: 03/11/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0369
Location:	WASHINGTON TOWNSHIP PD		
Calibration File No.:	00008	Calib. Date:	10/18/2005
Certification File No.:	00009	Cert. Date:	10/18/2005
Linearity File No.:	00010	Lin. Date:	10/18/2005
Solution File No.:	00067	Soln. Date:	02/20/2006
Sequential File No.:	00067	File Date:	02/20/2006
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0329
Solution Control Lot:	05F019	Expires:	06/03/2007
		Bottle No.:	0633

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:10S		
Control 1 EC	0.100%	15:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:11S		
Control 2 EC	0.097%	15:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12S		
Control 3 EC	0.097%	15:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOTTERI

First Name: DOMINICK

MI: A

Signature: _____



Badge No.: 204

Date: 02/20/2006

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Calib. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00061	Lin. No.: 00001
Sequential File No.:	00061	Soln. No.: 00006
		File Date: 02/09/2006
Calibrating Unit:	WET	Model No.: 4
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0632

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:10S		
Control 1 EC	0.099%	15:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12S		
Control 2 EC	0.098%	15:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13S		
Control 3 EC	0.097%	15:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: 

Badge No.: 513

Date: 02/09/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0369
Location:	WASHINGTON TOWNSHIP PD	
Calibration File No.:	00008	Calib. Date: 10/18/2005
Certification File No.:	00009	Cert. No.: 00003
Linearity File No.:	00010	Cert. Date: 10/18/2005
Solution File No.:	00046	Lin. No.: 00001
Sequential File No.:	00046	Soln. Date: 01/12/2006
		Soln. No.: 00005
Calibrating Unit:	WET	Model No.: 4
Control Solution %:	0.100%	Serial No.: DDWJ S3-0329
Solution Control Lot:	05F019	Expires: 06/03/2007
		Bottle No.: 0627

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:17S		
Control 1 EC	0.099%	15:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18S		
Control 2 EC	0.099%	15:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20S		
Control 3 EC	0.097%	15:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BOCCANFUSO

First Name: BRYAN

MI: S

Signature: _____



Badge No.: 513

Date: 01/12/2006